

EMBLEM™ MRI S-ICD SYSTEM

Subcutaneous Implantable Defibrillator

System Specifications

The EMBLEM MRI S-ICD is the second device in the EMBLEM S-ICD family and builds on previous size, longevity and remote patient management enhancements. Like transvenous ICDs, the EMBLEM MRI S-ICD System utilizes a pulse generator capable of delivering life-saving therapy. Unlike traditional ICDs, the EMBLEM MRI S-ICD System leaves the heart and vasculature untouched, avoiding potential complications associated with transvenous leads.

The EMBLEM MRI S-ICD has been tested and approved for use in the MR environment when the conditions of use are met. It contains a separate MRI mode with a timer that will automatically return the device to programmed settings. AF Monitor™ has also been added. This is a tool designed to assist in the detection of new onset, silent, or the progression of AF through R-R variability. The new SMART Pass filter is designed to reduce cardiac over-sensing and bench testing has demonstrated a > 40% reduction in inappropriate therapy.

Pulse Generator Specifications^{1,2}

Mechanical Specifications

Model Number	A219
Size (W x H x D)	83.1 x 69.1 x 12.7 mm
Mass	130g
Volume	59.5 cc (cm ³)
Longevity	7.3 years*
Battery	Boston Scientific Li/MnO ₂
Device C-Code	C1722



NEW ImageReady™ MR-Conditional Technology

Compatible Electrodes	3400, 3401, 3501
Magnet Strength	1.5T
Specific Absorption Rate (SAR) limits for the entire active scan (Normal Operating Mode)	<ul style="list-style-type: none"> • Whole body averaged, ≤ 2.0 watts/kilogram (W/kg) • Head, ≤ 3.2 W/kg
There are no anatomical exclusion zones or time restrictions.	

Programmable Parameters

Shock Zone	170 bpm - 250 bpm (steps of 10 bpm)
Conditional Shock Zone	Off, On 170 bpm - 240 bpm (minimum 10 bpm less than Shock Zone)
S-ICD System Therapy	Off, On
Post-shock pacing	Off, On (50 ppm, max 30 sec, demand-based)
Induction capability	1-10 sec (50 Hz/200 mA)
Delivered Energy	80J biphasic (only programmable during manual shock and induction test: 10J - 80J, steps of 5J)
Shocks per episode	Maximum of 5 shocks

Diagnostic Tools

NEW AF Monitor	Information Provided: <ul style="list-style-type: none"> • Number of days with measured AF in the last 90 days • Estimate of measured AF in the last 90 days (%) Performance: Sensitivity ≥ 87 % Positive Predictive Value ≥ 90 %
Episode storage	S-ECG storage for over 40 arrhythmic events (treated & untreated)
Other data	Electrode impedance System status (remaining battery life, patient alerts, etc.) Date and time stamp


* Longevity projections and the associated energy consumption is based on bench testing only.

1. EMBLEM MRI S-ICD User's Manual 359480-001 EN US 2015-11

2. MRI Technical Guide 359474-001 EN US 2015-11

EMBLEM™ MRI S-ICD SYSTEM

Subcutaneous Electrode Specification



Specifications

Model Number	3501
Type	Tripolar
Length	45 cm
Distal tip size (Diameter)	11.5 Fr / 3.84 mm
Coil size (Diameter)	9 Fr / 3 mm
Electrode shaft size (Diameter)	7 Fr / 2.33 mm
Sensing surface area	
Distal	36 mm ²
Proximal	46 mm ²
Sensing location	
Distal	At tip
Proximal	120 mm from tip
Defibrillation surface area	750 mm ²
Defibrillation location	20 - 100 mm from tip
Materials	
Insulation	Polycarbonate polyurethane
Electrodes	MP35N
Conductors	MP35N
Connector pin	MP35N
Integrated Suture Sleeve	Radiopaque White Silicone
Slit Suture Sleeve	Silicone
Electrode C-Code	C1896



**Boston
Scientific**
Advancing science for life™

CAUTION! Graphic pictures might not represent the actual size of the displayed products. All trademarks are the property of their respective owners. Law restricts this device to sale by or on the order of a physician. Indications, contraindications, warnings and instructions for use can be found in the product labeling supplied with each device. Information for the use only in countries with applicable health authority product registrations. Information not intended for use or distribution in France.

CRM-455415-AB MAR2017 Printed in Germany by medicalvision.

www.bostonscientific.eu

© 2017 Boston Scientific Corporation or its affiliates. All rights reserved.
DINCRM0956EA